



---

## CERTIFICATE OF ANALYSIS

**Product: 199-EBSS medium – 1 X Concentrated  
With Sodium Bicarbonate (2.2 g/L)**

**Lot # 230.817TB1**

**Catalog # BR30006-05**

**Manufacture Date: Ago/18/2017  
Expiration Date: Ago/18/2018**

<b>Test</b>	<b>Specification</b>	<b>Units</b>	<b>Results</b>
<b>Appearance</b>	<b>Clear, transparent</b>		<b>Clear, transparent</b>
<b>pH w/ NaHCO<sub>3</sub></b>	<b>7.2 +/- 0.2</b>		<b>7.21</b>
<b>Bacteria and yeast</b>	<b>Negative</b>		<b>Negative</b>
<b>Endotoxins</b>	<b>≤ 10.0</b>	<b>EU/mL</b>	<b>≤ 5</b>
<b><i>Mycoplasma</i>(qPCR)</b>	<b>Negative</b>		<b>Negative</b>
<b>Osmolality w/NaHCO<sub>3</sub></b>	<b>285 +/- 5%</b>	<b>mOsm/Kg H<sub>2</sub>O</b>	<b>290</b>

Electronically signed

**Rua Pasadena, 235 – Pq. Empresarial San José  
Cotia – SP / Brazil - 06715-864  
Phone-Fax: 55 11 46148070**